

## Certificate of analyses

**Product:** EnzyPlus Citric Acid  
**Art. No:** EZA 785  
**Batch No:** 400000  
**Expiry date:** July 2005

Diffchamb certifies that this batch has been approved by the Quality Control Department and that it conforms to the following specifications:

### Specific activity

Citrate Lyase  $\geq 50$  U/ml

### pH value

Vial 1 7.4 - 8.0

### Solubilisation time

Vial 1  $\leq 3$  min

Vial 2  $\leq 3$  min

### Performance test

target values  $\pm 5\%$

Date: 2004-06-16



**Jörgen Plymouth**  
Quality Manager  
Diffchamb AB

Diffchamb AB

F O Petersons Gata 32  
SE-421 31 Västra Frölunda  
[www.diffchamb.com](http://www.diffchamb.com)

Telephone +46-(0)31 742 33 50  
Telefax +46-(0)31 58 33 70  
E-mail [diffchamb@diffchamb.se](mailto:diffchamb@diffchamb.se)

Org.No. 556314-3303  
VAT No. SE 556314330301

Bankgiro 896-5576  
Postgiro 85 01 54-6



ISO9001  
Certificate  
No 11184