

## Certificate of analyses

**Product:** EnzyPlus L-Malic Acid  
**Art. No:** EZA 786  
**Batch No:** 400025  
**Expiry date:** October 2005

Diffchamb certifies that this batch has been approved by the Quality Control Department and that it conforms to the following specifications:

### Specific activity

L-MDH  $\geq$  6600 U/ml  
GOT  $\geq$  400 U/ml

### pH value

Vial 1 9.8 – 10.2

### Solubilisation time

Vial 2  $\leq$  3 min

### Performance test

target values  $\pm$  5%

Date: 2004-07-12



**Jörgen Plymouth**  
Quality Manager  
Diffchamb AB

Diffchamb AB

F O Petersons Gata 32  
SE-421 31 Västra Frölunda  
[www.diffchamb.com](http://www.diffchamb.com)

Telephone +46-(0)31 742 33 50  
Telefax +46-(0)31 58 33 70  
E-mail [diffchamb@diffchamb.se](mailto:diffchamb@diffchamb.se)

Org.No. 556314-3303  
VAT No. SE 556314330301

Bankgiro 896-5576  
Postgiro 85 01 54-6



ISO9001  
Certificate  
No 11184